



HPT Sensor Calibration Certificate

Part Number: 210091
 Maximum Pressure (psi): 100

Serial Number: XD31038A
 Calibration date: May 16, 2018
 Technician: SD

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.947E-1	-9.588E-1

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.660	15.666	14.624	0.036
20 PSI	20.130	21.184	20.113	0.017
30 PSI	30.170	31.305	30.180	-0.010
40 PSI	40.130	41.340	40.162	-0.032
50 PSI	50.480	51.755	50.522	-0.042
60 PSI	60.410	61.726	60.440	-0.030
70 PSI	70.320	71.657	70.318	0.002
80 PSI	80.670	82.015	80.622	0.048